

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111025\
 Data File : VN088236.D
 Acq On : 10 Nov 2025 13:07
 Operator : JC\MD
 Sample : VN1110MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1110MBL01

Quant Time: Nov 11 03:50:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	214241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	480036	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	462720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	223479	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	222767	60.131	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.260%
35) Dibromofluoromethane	8.147	113	171484	53.175	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.340%
50) Toluene-d8	10.547	98	611516	49.212	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.420%
62) 4-Bromofluorobenzene	12.829	95	233423	53.190	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111025\
Data File : VN088236.D
Acq On : 10 Nov 2025 13:07
Operator : JC\MD
Sample : VN1110MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1110MBL01

Quant Time: Nov 11 03:50:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

