

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111025\
 Data File : VN088237.D
 Acq On : 10 Nov 2025 13:27
 Operator : JC\MD
 Sample : VN1110WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1110WBL01

Quant Time: Nov 11 03:51:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	250665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	571291	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	544768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	267720	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	252918	58.350	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.700%
35) Dibromofluoromethane	8.147	113	185521	48.338	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.680%
50) Toluene-d8	10.547	98	665647	45.012	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.020%
62) 4-Bromofluorobenzene	12.829	95	251225	48.103	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111025\
Data File : VN088237.D
Acq On : 10 Nov 2025 13:27
Operator : JC\MD
Sample : VN1110WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1110WBL01

Quant Time: Nov 11 03:51:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

