

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111119\
 Data File : VN059172.D
 Acq On : 11 Nov 2019 19:20
 Operator : JC/SP
 Sample : K5760-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 3425

Quant Time: Nov 12 04:08:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	405396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	639462	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	563977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	200895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	273263	54.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.68%	
35) Dibromofluoromethane	7.57	113	206767	54.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.30%	
50) Toluene-d8	10.08	98	786179	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
62) 4-Bromofluorobenzene	12.40	95	249206	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
Target Compounds						
18) Methyl Acetate	4.31	43	8305	1.350	ug/l #	64
20) Methylene Chloride	4.52	84	6207	1.375	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111119\
Data File : VN059172.D
Acq On : 11 Nov 2019 19:20
Operator : JC/SP
Sample : K5760-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
3425

Quant Time: Nov 12 04:08:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

