

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111120\
 Data File : VN064546.D
 Acq On : 11 Nov 2020 11:22
 Operator : JC/MD
 Sample : VN1111WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1111WBL01

Quant Time: Nov 12 01:56:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	246099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	494949	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	511966	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	184284	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	239845	55.95	ug/l	0.00
Spiked Amount			Recovery	=	111.90%	
35) Dibromofluoromethane	7.55	113	178489	52.97	ug/l	0.00
Spiked Amount			Recovery	=	105.94%	
50) Toluene-d8	10.06	98	638217	49.00	ug/l	0.00
Spiked Amount			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	12.38	95	234841	47.68	ug/l	0.00
Spiked Amount			Recovery	=	95.36%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111120\
Data File : VN064546.D
Acq On : 11 Nov 2020 11:22
Operator : JC/MD
Sample : VN1111WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1111WBL01

Quant Time: Nov 12 01:56:19 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

