

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111120\
 Data File : VN064556.D
 Acq On : 11 Nov 2020 15:37
 Operator : JC/MD
 Sample : PB132934TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132934TB

Quant Time: Nov 12 02:27:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	238877	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	484646	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	538175	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	194249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235368	56.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.14%	
35) Dibromofluoromethane	7.55	113	172520	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
50) Toluene-d8	10.06	98	659738	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	12.38	95	248980	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
Target Compounds						
16) Acetone	3.79	43	13872	8.752	ug/l	95
18) Methyl Acetate	4.29	43	5243	1.683	ug/l #	90
20) Methylene Chloride	4.50	84	5810	1.848	ug/l	86
43) Isopropyl Acetate	8.13	43	105146	12.672	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111120\
Data File : VN064556.D
Acq On : 11 Nov 2020 15:37
Operator : JC/MD
Sample : PB132934TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132934TB

Quant Time: Nov 12 02:27:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

