

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111120\
 Data File : VN064562.D
 Acq On : 11 Nov 2020 18:03
 Operator : JC/MD
 Sample : PB132934ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132934ZHE#06

Manual Integrations
 APPROVED

MMDadoda
 11/12/2020 9:12:54 AM

Quant Time: Nov 12 02:38:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	233509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	518727	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	550634	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	192347	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	233645	57.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.88%	
35) Dibromofluoromethane	7.55	113	175114	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	10.06	98	667450	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.38	95	249677	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
Target Compounds						
16) Acetone	3.79	43	9301m	6.003	ug/l	
18) Methyl Acetate	4.29	43	3799m	1.247	ug/l	
20) Methylene Chloride	4.51	84	8918	2.901	ug/l	92
43) Isopropyl Acetate	8.13	43	44749	5.039	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111120\
Data File : VN064562.D
Acq On : 11 Nov 2020 18:03
Operator : JC/MD
Sample : PB132934ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132934ZHE#06

Manual Integrations
APPROVED

MMDadoda
11/12/2020 9:12:54 AM

Quant Time: Nov 12 02:38:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

