

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111120\
 Data File : VN064565.D
 Acq On : 11 Nov 2020 19:15
 Operator : JC/MD
 Sample : PB132934ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132934ZHE#09

Manual Integrations
 APPROVED

MMDadoda
 11/12/2020 9:13:05 AM

Quant Time: Nov 12 02:44:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	227930	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	506560	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	540162	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	234491	59.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.12%	
35) Dibromofluoromethane	7.55	113	174597	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
50) Toluene-d8	10.06	98	657654	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	12.38	95	243590	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
Target Compounds						
16) Acetone	3.78	43	13160	8.702	ug/l	92
18) Methyl Acetate	4.29	43	7604m	2.558	ug/l	
20) Methylene Chloride	4.52	84	7681m	2.560	ug/l	
43) Isopropyl Acetate	8.13	43	102093	11.772	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111120\
Data File : VN064565.D
Acq On : 11 Nov 2020 19:15
Operator : JC/MD
Sample : PB132934ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132934ZHE#09

Manual Integrations
APPROVED

MMDadoda
11/12/2020 9:13:05 AM

Quant Time: Nov 12 02:44:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

