

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111120\
 Data File : VN064567.D
 Acq On : 11 Nov 2020 20:04
 Operator : JC/MD
 Sample : PB132934ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132934ZHE#11

Quant Time: Nov 12 02:47:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	235420	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	515278	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	547510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	193313	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	234793	57.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.52%	
35) Dibromofluoromethane	7.55	113	172648	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
50) Toluene-d8	10.06	98	670400	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	12.38	95	245241	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
Target Compounds						
16) Acetone	3.78	43	14226	9.107	ug/l	96
18) Methyl Acetate	4.28	43	6863	2.235	ug/l #	70
20) Methylene Chloride	4.51	84	7745	2.499	ug/l #	82
43) Isopropyl Acetate	8.13	43	101342	11.488	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111120\
Data File : VN064567.D
Acq On : 11 Nov 2020 20:04
Operator : JC/MD
Sample : PB132934ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132934ZHE#11

Quant Time: Nov 12 02:47:45 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

