

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
 Data File : VN069451.D
 Acq On : 11 Nov 2021 11:04
 Operator : JC/MD
 Sample : VN1111MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1111MBL01

Quant Time: Nov 12 05:15:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1260115	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2226117	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2131654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	767715	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	925192	50.637	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.280%			
35) Dibromofluoromethane	8.024	113	654588	49.028	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.060%			
50) Toluene-d8	10.443	98	2795909	53.245	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.480%			
62) 4-Bromofluorobenzene	12.733	95	941573	48.810	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 97.620%			
Target Compounds						Qvalue
93) 1,2,4-Trichlorobenzene	15.265	180	11429	4.000	ug/l	97
95) Naphthalene	15.509	128	48861	5.757	ug/l #	93
96) 1,2,3-Trichlorobenzene	15.700	180	13719	4.281	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
Data File : VN069451.D
Acq On : 11 Nov 2021 11:04
Operator : JC/MD
Sample : VN1111MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1111MBL01

Quant Time: Nov 12 05:15:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

