

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
 Data File : VN069455.D
 Acq On : 11 Nov 2021 14:27
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Nov 12 05:16:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1229535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2182422	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2143596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	810232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	919094	51.555	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.100%
35) Dibromofluoromethane	8.024	113	645149	49.288	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.580%
50) Toluene-d8	10.443	98	2785549	54.110	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.220%
62) 4-Bromofluorobenzene	12.734	95	965174	51.036	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.080%
Target Compounds						
						Qvalue
93) 1,2,4-Trichlorobenzene	15.265	180	8210	3.717	ug/l	91
95) Naphthalene	15.509	128	35094	5.340	ug/l #	94
96) 1,2,3-Trichlorobenzene	15.702	180	9577	3.905	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
Data File : VN069455.D
Acq On : 11 Nov 2021 14:27
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Nov 12 05:16:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

