

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
 Data File : VN069458.D
 Acq On : 11 Nov 2021 17:05
 Operator : JC/MD
 Sample : M4660-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FM655

Quant Time: Nov 12 05:16:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1182090	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	2114591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2049273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	725465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	906525	52.890	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.780%
35) Dibromofluoromethane	8.024	113	626911	49.431	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.860%
50) Toluene-d8	10.443	98	2700951	54.149	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.300%
62) 4-Bromofluorobenzene	12.733	95	900001	49.116	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.240%
Target Compounds						
						Qvalue
3) Chloromethane	2.306	50	3351	0.229	ug/l	97
16) Acetone	4.299	43	10973	1.187	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
Data File : VN069458.D
Acq On : 11 Nov 2021 17:05
Operator : JC/MD
Sample : M4660-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FM655

Quant Time: Nov 12 05:16:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

