

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
 Data File : VN069459.D
 Acq On : 11 Nov 2021 17:38
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 12 05:17:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.534	0.513	3.9	88	0.00
3 P	Chloromethane	0.619	0.553	10.7	90	0.00
4 C	Vinyl Chloride	0.620	0.575	7.3#	89	0.00
5 T	Bromomethane	0.403	0.349	13.4	89	0.00
6 T	Chloroethane	0.419	0.389	7.2	91	0.00
7 T	Trichlorofluoromethane	0.959	0.901	6.0	90	0.00
8 T	Diethyl Ether	0.310	0.288	7.1	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.452	0.432	4.4	93	0.00
10 T	Methyl Iodide	0.554	0.534	3.6	92	0.00
11 T	Tert butyl alcohol	0.131	0.114	13.0	86	0.00
12 CM	1,1-Dichloroethene	0.405	0.389	4.0#	92	0.00
13 T	Acrolein	0.021	0.018	14.3	95	0.00
14 T	Allyl chloride	0.896	0.848	5.4	91	0.00
15 T	Acrylonitrile	0.335	0.325	3.0	93	0.00
16 T	Acetone	0.391	0.335	14.3	82	0.00
17 T	Carbon Disulfide	0.958	0.817	14.7	82	0.00
18 T	Methyl Acetate	1.272	1.158	9.0	90	0.00
19 T	Methyl tert-butyl Ether	1.715	1.691	1.4	94	0.00
20 T	Methylene Chloride	0.555	0.500	9.9	92	0.00
21 T	trans-1,2-Dichloroethene	0.453	0.426	6.0	92	0.00
22 T	Diisopropyl ether	1.930	1.929	0.1	95	0.00
23 T	Vinyl Acetate	1.493	1.503	-0.7	93	0.00
24 P	1,1-Dichloroethane	0.983	0.958	2.5	95	0.00
25 T	2-Butanone	0.530	0.499	5.8	89	0.00
26 T	2,2-Dichloropropane	0.769	0.737	4.2	89	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.558	0.5	95	0.00
28 T	Bromochloromethane	0.503	0.515	-2.4	102	0.00
29 T	Tetrahydrofuran	0.325	0.318	2.2	90	0.00
30 C	Chloroform	1.008	1.015	-0.7#	97	0.00
31 T	Cyclohexane	0.990	0.887	10.4	92	0.00
32 T	1,1,1-Trichloroethane	0.841	0.847	-0.7	94	0.00
33 S	1,2-Dichloroethane-d4	0.725	0.702	3.2	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.300	0.297	1.0	100	0.00
36 T	1,1-Dichloropropene	0.421	0.421	0.0	93	0.00
37 T	Ethyl Acetate	0.569	0.583	-2.5	93	0.00
38 T	Carbon Tetrachloride	0.406	0.409	-0.7	93	0.00
39 T	Methylcyclohexane	0.523	0.518	1.0	90	0.00
40 TM	Benzene	1.299	1.301	-0.2	95	0.00
41 T	Methacrylonitrile	0.294	0.289	1.7	97	0.00
42 TM	1,2-Dichloroethane	0.508	0.512	-0.8	96	0.00
43 T	Isopropyl Acetate	0.913	0.909	0.4	92	0.00
44 TM	Trichloroethene	0.311	0.312	-0.3	95	0.00
45 C	1,2-Dichloropropane	0.347	0.354	-2.0#	95	0.00
46 T	Dibromomethane	0.232	0.234	-0.9	95	0.00
47 T	Bromodichloromethane	0.434	0.454	-4.6	94	0.00
48 T	Methyl methacrylate	0.430	0.420	2.3	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
 Data File : VN069459.D
 Acq On : 11 Nov 2021 17:38
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 12 05:17:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	92	0.00
50 S	Toluene-d8	1.179	1.201	-1.9	99	0.00
51 T	4-Methyl-2-Pentanone	0.590	0.615	-4.2	93	0.00
52 CM	Toluene	0.809	0.844	-4.3#	95	0.00
53 T	t-1,3-Dichloropropene	0.453	0.484	-6.8	93	0.00
54 T	cis-1,3-Dichloropropene	0.495	0.524	-5.9	94	0.00
55 T	1,1,2-Trichloroethane	0.326	0.347	-6.4	98	0.00
56 T	Ethyl methacrylate	0.540	0.573	-6.1	95	0.00
57 T	1,3-Dichloropropane	0.573	0.594	-3.7	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.132	-5.6	95	0.00
59 T	2-Hexanone	0.439	0.448	-2.1	90	0.00
60 T	Dibromochloromethane	0.305	0.336	-10.2	96	0.00
61 T	1,2-Dibromoethane	0.330	0.353	-7.0	96	0.00
62 S	4-Bromofluorobenzene	0.433	0.453	-4.6	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.302	0.292	3.3	94	0.00
65 PM	Chlorobenzene	0.931	0.927	0.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.323	0.335	-3.7	95	0.00
67 C	Ethyl Benzene	1.716	1.740	-1.4#	96	0.00
68 T	m/p-Xylenes	0.631	0.641	-1.6	95	0.00
69 T	o-Xylene	0.629	0.647	-2.9	96	0.00
70 T	Styrene	1.011	1.062	-5.0	96	0.00
71 P	Bromoform	0.230	0.245	-6.5	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.981	3.767	5.4	96	0.00
74 T	N-amyl acetate	1.687	1.640	2.8	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.343	1.255	6.6	96	0.00
76 T	1,2,3-Trichloropropane	1.154	1.196	-3.6	104	0.00
77 T	Bromobenzene	0.910	0.861	5.4	97	0.00
78 T	n-propylbenzene	4.480	4.331	3.3	96	0.00
79 T	2-Chlorotoluene	2.774	2.603	6.2	96	0.00
80 T	1,3,5-Trimethylbenzene	3.218	3.085	4.1	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.344	0.318	7.6	91	0.00
82 T	4-Chlorotoluene	2.644	2.507	5.2	96	0.00
83 T	tert-Butylbenzene	2.820	2.728	3.3	96	0.00
84 T	1,2,4-Trimethylbenzene	3.107	3.047	1.9	96	0.00
85 T	sec-Butylbenzene	3.940	3.856	2.1	96	0.00
86 T	p-Isopropyltoluene	3.128	3.121	0.2	96	0.00
87 T	1,3-Dichlorobenzene	1.579	1.533	2.9	96	0.00
88 T	1,4-Dichlorobenzene	1.549	1.483	4.3	96	0.00
89 T	n-Butylbenzene	2.608	2.586	0.8	94	0.00
90 T	Hexachloroethane	0.513	0.507	1.2	93	0.00
91 T	1,2-Dichlorobenzene	1.502	1.492	0.7	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.217	0.216	0.5	92	0.00
93 T	1,2,4-Trichlorobenzene	0.662	0.685	-3.5	93	0.00
94 T	Hexachlorobutadiene	0.390	0.385	1.3	94	0.00
95 T	Naphthalene	1.775	1.813	-2.1	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111121\
Data File : VN069459.D
Acq On : 11 Nov 2021 17:38
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 12 05:17:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.634	0.667	-5.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6