

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111122\
 Data File : VN075249.D
 Acq On : 11 Nov 2022 15:41
 Operator : JC\MD
 Sample : N5582-09 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CLI-0669

Quant Time: Nov 14 03:46:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	103645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	197329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	170158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	74429	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	74021	53.005	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.000%
35) Dibromofluoromethane	8.177	113	64624	55.786	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.580%
50) Toluene-d8	10.571	98	220432	49.046	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.100%
62) 4-Bromofluorobenzene	12.847	95	75005	47.508	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111122\
Data File : VN075249.D
Acq On : 11 Nov 2022 15:41
Operator : JC\MD
Sample : N5582-09 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CLI-0669

Quant Time: Nov 14 03:46:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

