

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111122\
 Data File : VN075250.D
 Acq On : 11 Nov 2022 16:05
 Operator : JC\MD
 Sample : N5582-10 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CLI-0667

Quant Time: Nov 14 03:46:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	103199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	192613	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	168682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72124	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	74363	53.480	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.960%	
35) Dibromofluoromethane	8.171	113	63223	55.913	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.820%	
50) Toluene-d8	10.571	98	218004	49.694	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.380%	
62) 4-Bromofluorobenzene	12.853	95	73595	47.756	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111122\
Data File : VN075250.D
Acq On : 11 Nov 2022 16:05
Operator : JC\MD
Sample : N5582-10 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CLI-0667

Quant Time: Nov 14 03:46:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

