

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111122\
 Data File : VN075255.D
 Acq On : 11 Nov 2022 18:05
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 14 03:49:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	130	0.00
2 T	Dichlorodifluoromethane	0.650	0.600	7.7	123	0.00
3 P	Chloromethane	0.691	0.528	23.6	104	0.00
4 C	Vinyl Chloride	0.986	0.733	25.7#	98	0.00
5 T	Bromomethane	0.808	0.671	17.0	112	0.00
6 T	Chloroethane	0.707	0.524	25.9#	101	0.00
7 T	Trichlorofluoromethane	1.007	0.977	3.0	130	0.00
8 T	Diethyl Ether	0.367	0.356	3.0	128	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.556	0.542	2.5	131	0.00
10 T	Methyl Iodide	0.828	0.860	-3.9	143	0.00
11 T	Tert butyl alcohol	0.157	0.128	18.5	116	0.00
12 CM	1,1-Dichloroethene	0.530	0.512	3.4#	130	0.00
13 T	Acrolein	0.129	0.109	15.5	116	0.00
14 T	Allyl chloride	0.774	0.690	10.9	116	0.00
15 T	Acrylonitrile	0.332	0.302	9.0	123	0.00
16 T	Acetone	0.285	0.245	14.0	122	0.00
17 T	Carbon Disulfide	1.333	1.181	11.4	117	0.00
18 T	Methyl Acetate	0.932	0.836	10.3	122	0.00
19 T	Methyl tert-butyl Ether	1.970	1.848	6.2	124	0.00
20 T	Methylene Chloride	0.636	0.583	8.3	126	0.00
21 T	trans-1,2-Dichloroethene	0.580	0.561	3.3	131	0.00
22 T	Diisopropyl ether	1.708	1.504	11.9	116	0.00
23 T	Vinyl Acetate	1.427	1.242	13.0	113	0.00
24 P	1,1-Dichloroethane	1.080	0.982	9.1	122	0.00
25 T	2-Butanone	0.449	0.388	13.6	118	0.00
26 T	2,2-Dichloropropane	0.994	0.904	9.1	122	0.00
27 T	cis-1,2-Dichloroethene	0.693	0.677	2.3	130	0.00
28 T	Bromochloromethane	0.455	0.363	20.2	105	0.00
29 T	Tetrahydrofuran	0.290	0.255	12.1	116	0.00
30 C	Chloroform	1.173	1.084	7.6#	122	0.00
31 T	Cyclohexane	1.002	0.858	14.4	116	0.00
32 T	1,1,1-Trichloroethane	1.065	1.016	4.6	125	0.00
33 S	1,2-Dichloroethane-d4	0.674	0.541	19.7	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	118	0.00
35 S	Dibromofluoromethane	0.294	0.291	1.0	121	0.00
36 T	1,1-Dichloropropene	0.454	0.464	-2.2	122	0.00
37 T	Ethyl Acetate	0.498	0.456	8.4	117	0.00
38 T	Carbon Tetrachloride	0.500	0.540	-8.0	127	0.00
39 T	Methylcyclohexane	0.600	0.621	-3.5	123	0.00
40 TM	Benzene	1.391	1.405	-1.0	122	0.00
41 T	Methacrylonitrile	0.239	0.234	2.1	118	0.00
42 TM	1,2-Dichloroethane	0.506	0.472	6.7	113	0.00
43 T	Isopropyl Acetate	0.773	0.717	7.2	111	0.00
44 TM	Trichloroethene	0.349	0.390	-11.7	138	0.00
45 C	1,2-Dichloropropane	0.331	0.340	-2.7#	121	0.00
46 T	Dibromomethane	0.258	0.263	-1.9	125	0.00
47 T	Bromodichloromethane	0.500	0.492	1.6	119	0.00
48 T	Methyl methacrylate	0.355	0.341	3.9	120	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111122\
 Data File : VN075255.D
 Acq On : 11 Nov 2022 18:05
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 14 03:49:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	130	0.00
50 S	Toluene-d8	1.139	1.074	5.7	114	0.00
51 T	4-Methyl-2-Pentanone	0.499	0.467	6.4	114	0.00
52 CM	Toluene	0.930	0.966	-3.9#	119	0.00
53 T	t-1,3-Dichloropropene	0.527	0.539	-2.3	116	0.00
54 T	cis-1,3-Dichloropropene	0.559	0.575	-2.9	120	0.00
55 T	1,1,2-Trichloroethane	0.369	0.369	0.0	118	0.00
56 T	Ethyl methacrylate	0.563	0.573	-1.8	117	0.00
57 T	1,3-Dichloropropane	0.614	0.615	-0.2	119	0.00
58 T	2-Chloroethyl Vinyl ether	0.217	0.212	2.3	110	0.00
59 T	2-Hexanone	0.379	0.353	6.9	113	0.00
60 T	Dibromochloromethane	0.389	0.427	-9.8	125	0.00
61 T	1,2-Dibromoethane	0.378	0.411	-8.7	125	0.00
62 S	4-Bromofluorobenzene	0.400	0.407	-1.7	116	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	117	0.00
64 T	Tetrachloroethene	0.307	0.377	-22.8	152#	0.00
65 PM	Chlorobenzene	1.067	1.115	-4.5	126	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.425	-7.3	127	0.00
67 C	Ethyl Benzene	1.920	1.972	-2.7#	119	0.00
68 T	m/p-Xylenes	0.770	0.808	-4.9	119	0.00
69 T	o-Xylene	0.762	0.811	-6.4	124	0.00
70 T	Styrene	1.215	1.311	-7.9	122	0.00
71 P	Bromoform	0.296	0.356	-20.3	138	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
73 T	Isopropylbenzene	4.014	3.841	4.3	120	0.00
74 T	N-amyl acetate	1.485	1.289	13.2	112	0.00
75 P	1,1,2,2-Tetrachloroethane	1.309	1.188	9.2	120	0.00
76 T	1,2,3-Trichloropropane	1.062	0.998	6.0	113	0.00
77 T	Bromobenzene	0.884	0.941	-6.4	137	0.00
78 T	n-propylbenzene	4.507	4.402	2.3	121	0.00
79 T	2-Chlorotoluene	2.754	2.600	5.6	118	0.00
80 T	1,3,5-Trimethylbenzene	3.371	3.310	1.8	120	0.00
81 T	trans-1,4-Dichloro-2-butene	0.414	0.358	13.5	107	0.00
82 T	4-Chlorotoluene	2.754	2.600	5.6	118	0.00
83 T	tert-Butylbenzene	2.996	2.957	1.3	123	0.00
84 T	1,2,4-Trimethylbenzene	3.411	3.348	1.8	120	0.00
85 T	sec-Butylbenzene	4.191	4.254	-1.5	123	0.00
86 T	p-Isopropyltoluene	3.461	3.651	-5.5	126	0.00
87 T	1,3-Dichlorobenzene	1.731	1.825	-5.4	130	0.00
88 T	1,4-Dichlorobenzene	1.713	1.795	-4.8	130	0.00
89 T	n-Butylbenzene	2.787	3.056	-9.7	128	0.00
90 T	Hexachloroethane	0.635	0.626	1.4	120	0.00
91 T	1,2-Dichlorobenzene	1.720	1.830	-6.4	132	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.259	4.4	120	0.00
93 T	1,2,4-Trichlorobenzene	0.740	1.032	-39.5#	164#	0.00
94 T	Hexachlorobutadiene	0.361	0.497	-37.7#	177#	0.00
95 T	Naphthalene	2.847	3.476	-22.1	144	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111122\
Data File : VN075255.D
Acq On : 11 Nov 2022 18:05
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 14 03:49:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.736	0.990	-34.5#	162#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6