

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084769.D
 Acq On : 11 Nov 2024 14:06
 Operator : JC\MD
 Sample : VN1111WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Nov 12 00:30:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	168331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	276055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	247853	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129321	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113397	46.641	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.280%
35) Dibromofluoromethane	8.165	113	90978	48.689	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.380%
50) Toluene-d8	10.565	98	333596	48.465	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.940%
62) 4-Bromofluorobenzene	12.847	95	129173	50.214	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.420%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.118	85	30960	15.999	ug/l	92
3) Chloromethane	2.359	50	34232	13.222	ug/l	92
4) Vinyl Chloride	2.512	62	33565	16.127	ug/l	98
5) Bromomethane	2.901	94	19026	17.253	ug/l	99
6) Chloroethane	3.071	64	22580	16.594	ug/l	95
7) Trichlorofluoromethane	3.465	101	59707	17.415	ug/l	98
8) Diethyl Ether	3.953	74	20638	17.064	ug/l	91
9) 1,1,2-Trichlorotrifluo...	4.347	101	34692	17.907	ug/l	97
10) Methyl Iodide	4.577	142	45547	17.591	ug/l	98
11) Tert butyl alcohol	5.536	59	28762	89.623	ug/l	98
12) 1,1-Dichloroethene	4.324	96	30991	16.167	ug/l	97
13) Acrolein	4.171	56	23281	72.750	ug/l	99
14) Allyl chloride	5.012	41	50477	16.237	ug/l	93
15) Acrylonitrile	5.712	53	81152	87.349	ug/l	97
16) Acetone	4.430	43	66816	93.351	ug/l	99
17) Carbon Disulfide	4.700	76	83488	14.143	ug/l	99
18) Methyl Acetate	5.024	43	43723	14.184	ug/l	99
19) Methyl tert-butyl Ether	5.783	73	105991	18.039	ug/l	95
20) Methylene Chloride	5.265	84	37401	17.491	ug/l	92
21) trans-1,2-Dichloroethene	5.777	96	31467	15.986	ug/l	99
22) Diisopropyl ether	6.665	45	108244	17.523	ug/l	96
23) Vinyl Acetate	6.594	43	367953	86.376	ug/l	99
24) 1,1-Dichloroethane	6.559	63	65162	17.570	ug/l	96
25) 2-Butanone	7.483	43	98694	87.011	ug/l	92
26) 2,2-Dichloropropane	7.488	77	58631	18.164	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	39661	17.256	ug/l	95
28) Bromochloromethane	7.812	49	33725	22.825	ug/l	88
29) Tetrahydrofuran	7.841	42	64535	85.762	ug/l	93
30) Chloroform	7.959	83	67595	17.907	ug/l	99
31) Cyclohexane	8.253	56	55130	16.424	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	61402	17.872	ug/l	97
36) 1,1-Dichloropropene	8.365	75	44607	17.213	ug/l	99
37) Ethyl Acetate	7.559	43	41478	17.641	ug/l	98
38) Carbon Tetrachloride	8.359	117	54755	18.903	ug/l	97
39) Methylcyclohexane	9.600	83	47024	17.837	ug/l	97
40) Benzene	8.600	78	146291	17.578	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084769.D
 Acq On : 11 Nov 2024 14:06
 Operator : JC\MD
 Sample : VN1111WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Nov 12 00:30:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	22119	17.494	ug/l	91
42) 1,2-Dichloroethane	8.665	62	48432	17.969	ug/l	100
43) Isopropyl Acetate	8.688	43	73071	16.169	ug/l	96
44) Trichloroethene	9.353	130	35450	18.470	ug/l	93
45) 1,2-Dichloropropane	9.618	63	36769	18.834	ug/l	100
46) Dibromomethane	9.706	93	25326	18.384	ug/l	94
47) Bromodichloromethane	9.888	83	53909	18.565	ug/l #	99
48) Methyl methacrylate	9.677	41	31997	17.486	ug/l	95
49) 1,4-Dioxane	9.700	88	12215	376.695	ug/l #	82
51) 4-Methyl-2-Pentanone	10.447	43	210440	93.887	ug/l	98
52) Toluene	10.629	92	90899	18.675	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	52220	17.373	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	56090	17.631	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	34267	18.690	ug/l	95
56) Ethyl methacrylate	10.871	69	52072	19.635	ug/l #	89
57) 1,3-Dichloropropane	11.159	76	58151	18.480	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	107908	86.954	ug/l	95
59) 2-Hexanone	11.194	43	155905	95.560	ug/l	98
60) Dibromochloromethane	11.359	129	40894	19.570	ug/l	97
61) 1,2-Dibromoethane	11.465	107	34114	18.170	ug/l	100
64) Tetrachloroethene	11.100	164	29364	17.811	ug/l	94
65) Chlorobenzene	11.888	112	95953	17.305	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	36990	19.179	ug/l	95
67) Ethyl Benzene	11.965	91	161814	17.483	ug/l	99
68) m/p-Xylenes	12.071	106	125863	35.919	ug/l	99
69) o-Xylene	12.400	106	62468	19.057	ug/l	99
70) Styrene	12.412	104	104440	18.264	ug/l	98
71) Bromoform	12.576	173	27320	18.825	ug/l #	95
73) Isopropylbenzene	12.694	105	152918	16.874	ug/l	99
74) N-amyl acetate	12.494	43	61403	15.919	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.935	83	50456	16.680	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	65259	25.245	ug/l	57
77) Bromobenzene	12.982	156	39902	16.378	ug/l	92
78) n-propylbenzene	13.035	91	176633	16.633	ug/l	98
79) 2-Chlorotoluene	13.123	91	117842	16.982	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	134354	17.887	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	19551	17.943	ug/l	95
82) 4-Chlorotoluene	13.217	91	117134	16.040	ug/l	98
83) tert-Butylbenzene	13.435	119	110533	17.781	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	135654	17.499	ug/l	99
85) sec-Butylbenzene	13.617	105	148482	16.909	ug/l	99
86) p-Isopropyltoluene	13.729	119	126790	17.606	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	74392	15.646	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	74879	16.436	ug/l	99
89) n-Butylbenzene	14.053	91	102319	15.189	ug/l	98
90) Hexachloroethane	14.329	117	25970	16.171	ug/l	92
91) 1,2-Dichlorobenzene	14.106	146	73274	16.038	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	10323	16.845	ug/l	91
93) 1,2,4-Trichlorobenzene	15.394	180	35525	14.206	ug/l	95
94) Hexachlorobutadiene	15.500	225	17031	15.508	ug/l	97
95) Naphthalene	15.641	128	101990	13.626	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	33980	13.998	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
Data File : VN084769.D
Acq On : 11 Nov 2024 14:06
Operator : JC\MD
Sample : VN1111WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Nov 12 00:30:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 12 00:30:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 168331
Ion Ratio Lower Upper
168 100
99 63.2 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 15.999 ug/l
RT: 2.118 min Scan# 28
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 85 Resp: 30960
Ion Ratio Lower Upper
85 100
87 30.6 17.6 52.9









#3
Chloromethane
Concen: 13.222 ug/l
RT: 2.359 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 34232
Ion Ratio Lower Upper
50 100
52 28.7 26.8 40.2



#4
Vinyl Chloride
Concen: 16.127 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 62 Resp: 33565
Ion Ratio Lower Upper
62 100
64 31.7 26.2 39.4









#5
Bromomethane
Concen: 17.253 ug/l
RT: 2.901 min Scan# 11
Delta R.T. -0.052 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 19026
Ion Ratio Lower Upper
94 100
96 91.8 72.9 109.3



#6
Chloroethane
Concen: 16.594 ug/l
RT: 3.071 min Scan# 190
Delta R.T. -0.047 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 64 Resp: 22580
Ion Ratio Lower Upper
64 100
66 31.9 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 17.415 ug/l
 RT: 3.465 min Scan# 21
 Delta R.T. -0.030 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:101 Resp: 59707
 Ion Ratio Lower Upper
 101 100
 103 66.1 51.8 77.6



#8
 Diethyl Ether
 Concen: 17.064 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. -0.006 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06

Tgt Ion: 74 Resp: 20638
 Ion Ratio Lower Upper
 74 100
 45 86.2 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 17.907 ug/l
RT: 4.347 min Scan# 407
Delta R.T. -0.024 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101	Resp:	34692
Ion	Ratio	Lower Upper
101	100	
85	44.5	36.2 54.4
151	76.9	59.3 88.9



#10
Methyl Iodide
Concen: 17.591 ug/l
RT: 4.577 min Scan# 446
Delta R.T. -0.018 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion:142	Resp:	45547
Ion	Ratio	Lower Upper
142	100	
127	45.8	35.3 52.9
141	14.0	11.0 16.6









#11
Tert butyl alcohol
Concen: 89.623 ug/l
RT: 5.536 min Scan# 609
Delta R.T. 0.012 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 59 Resp: 28762
Ion Ratio Lower Upper
59 100
57 9.6 8.2 12.2



#12
1,1-Dichloroethene
Concen: 16.167 ug/l
RT: 4.324 min Scan# 403
Delta R.T. -0.018 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 96 Resp: 30991
Ion Ratio Lower Upper
96 100
61 156.8 128.6 192.8
98 67.7 51.7 77.5





#13

Acrolein

Concen: 72.750 ug/l

RT: 4.171 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 56 Resp: 23281

Ion Ratio Lower Upper

56 100

55 72.3 57.2 85.8



#14

Allyl chloride

Concen: 16.237 ug/l

RT: 5.012 min Scan# 520

Delta R.T. -0.012 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06




Tgt Ion: 41 Resp: 50477


Ion Ratio Lower Upper

41 100

39 65.4 59.0 88.6


76 36.4 29.8 44.6








#15
Acrylonitrile
Concen: 87.349 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 53 Resp: 81152
Ion Ratio Lower Upper
53 100
52 79.9 66.2 99.4
51 34.9 28.8 43.2



#16
Acetone
Concen: 93.351 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 43 Resp: 66816
Ion Ratio Lower Upper
43 100
58 29.8 24.4 36.6









#17
Carbon Disulfide
Concen: 14.143 ug/l
RT: 4.700 min Scan# 41
Delta R.T. -0.012 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 83488
Ion Ratio Lower Upper
76 100
78 8.8 7.3 10.9



#18
Methyl Acetate
Concen: 14.184 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 43 Resp: 43723
Ion Ratio Lower Upper
43 100
74 24.6 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 18.039 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 73 Resp: 105991
Ion Ratio Lower Upper
73 100
57 19.7 17.7 26.5



#20
Methylene Chloride
Concen: 17.491 ug/l
RT: 5.265 min Scan# 563
Delta R.T. -0.012 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 84 Resp: 37401
Ion Ratio Lower Upper
84 100
49 117.5 107.2 160.8
51 40.6 31.8 47.8
86 65.8 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 15.986 ug/l
RT: 5.777 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 31467
Ion Ratio Lower Upper
96 100
61 136.3 108.4 162.6
98 60.7 49.8 74.8



#22
Diisopropyl ether
Concen: 17.523 ug/l
RT: 6.665 min Scan# 801
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 45 Resp: 108244
Ion Ratio Lower Upper
45 100
43 48.7 42.3 63.5
87 25.8 21.8 32.6
59 12.2 9.6 14.4









#23
 Vinyl Acetate
 Concen: 86.376 ug/l
 RT: 6.594 min Scan# 790
 Delta R.T. -0.006 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 367953
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 17.570 ug/l
 RT: 6.559 min Scan# 783
 Delta R.T. -0.012 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06

Tgt Ion: 63 Resp: 65162
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.4 10.2
 100 5.8 2.0 5.8









#25
2-Butanone
Concen: 87.011 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 98694
Ion Ratio Lower Upper
43 100
72 31.1 21.4 32.2



#26
2,2-Dichloropropane
Concen: 18.164 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06


Tgt Ion: 77 Resp: 58631
Ion Ratio Lower Upper
77 100
97 22.2 10.8 32.3








#27
 cis-1,2-Dichloroethene
 Concen: 17.256 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. -0.006 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 96 Resp: 39661
 Ion Ratio Lower Upper
 96 100
 61 142.4 0.0 301.8
 98 64.4 0.0 128.2



#28
 Bromochloromethane
 Concen: 22.825 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06

Tgt Ion: 49 Resp: 33725
 Ion Ratio Lower Upper
 49 100
 129 1.1 0.0 3.6
 130 82.4 58.0 87.0





#29

Tetrahydrofuran

Concen: 85.762 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084769.D





Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	42	Resp:	64535
Ion	Ratio	Lower	Upper
42	100		
72	47.1	34.6	51.8
71	44.9	32.2	48.4



#30

Chloroform

Concen: 17.907 ug/l




RT: 7.959 min Scan# 1021


Delta R.T. -0.006 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Tgt Ion:	83	Resp:	67595
Ion	Ratio	Lower	Upper
83	100		
85	64.7	51.0	76.4









#31
Cyclohexane
Concen: 16.424 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 55130
Ion Ratio Lower Upper
56 100
69 35.0 23.8 35.6
84 88.4 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 17.872 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 97 Resp: 61402
Ion Ratio Lower Upper
97 100
99 64.5 51.2 76.8
61 42.9 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 46.641 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 113397

Ion Ratio Lower Upper

65 100

67 51.8 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06




Tgt Ion:114 Resp: 276055


Ion Ratio Lower Upper

114 100

63 20.0 0.0 43.8

88 16.1 0.0 31.6





#35

Dibromofluoromethane

Concen: 48.689 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:113 Resp: 90978

Ion Ratio Lower Upper

113 100

111 102.8 83.3 124.9

192 18.6 13.5 20.3



#36

1,1-Dichloropropene

Concen: 17.213 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. -0.006 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06




Tgt Ion: 75 Resp: 44607


Ion Ratio Lower Upper

75 100

110 34.4 16.9 50.7

77 31.6 24.7 37.1









#37
Ethyl Acetate
Concen: 17.641 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 41478
Ion Ratio Lower Upper
43 100
61 14.2 10.7 16.1
70 11.5 8.8 13.2



#38
Carbon Tetrachloride
Concen: 18.903 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 117 Resp: 54755
Ion Ratio Lower Upper
117 100
119 95.2 78.7 118.1
121 29.5 24.4 36.6





#39

Methylcyclohexane

Concen: 17.837 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 47024

Ion Ratio Lower Upper

83 100

55 80.7 62.8 94.2

98 51.7 38.8 58.2



#40

Benzene

Concen: 17.578 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. -0.006 min

Lab File: VN084769.D




Acq: 11 Nov 2024 14:06


Tgt Ion: 78 Resp: 146291

Ion Ratio Lower Upper

78 100

77 24.4 19.1 28.7









#41
Methacrylonitrile
Concen: 17.494 ug/l
RT: 7.771 min Scan# 911
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	41	Resp:	22119
Ion	Ratio	Lower	Upper
41	100		
39	57.3	43.3	64.9
67	81.9	56.1	84.1
52	33.2	24.5	36.7



#42
1,2-Dichloroethane
Concen: 17.969 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion:	62	Resp:	48432
Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	18.2









#43
Isopropyl Acetate
Concen: 16.169 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 73071
Ion Ratio Lower Upper
43 100
61 28.1 20.4 30.6
87 13.6 10.3 15.5



#44
Trichloroethene
Concen: 18.470 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion:130 Resp: 35450
Ion Ratio Lower Upper
130 100
95 94.0 0.0 203.0





#45

1,2-Dichloropropane

Concen: 18.834 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 36769

Ion Ratio Lower Upper

63 100

65 30.8 24.6 37.0



#46

Dibromomethane

Concen: 18.384 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06




Tgt Ion: 93 Resp: 25326


Ion Ratio Lower Upper

93 100

95 83.5 65.0 97.4

174 98.0 71.8 107.8





#47

Bromodichloromethane

Concen: 18.565 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN084769.D





Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 83	Resp: 53909
Ion Ratio	Lower Upper
83	100
85	64.8 51.5 77.3
127	8.9 5.7 8.5#



#48

Methyl methacrylate

Concen: 17.486 ug/l




RT: 9.677 min Scan# 1313


Delta R.T. 0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Tgt Ion: 41	Resp: 31997
Ion Ratio	Lower Upper
41	100
69	89.2 67.1 100.7
39	59.3 45.8 68.6





#49

1,4-Dioxane

Concen: 376.695 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 88 Resp: 12215

Ion Ratio Lower Upper

88 100

43 0.0 25.3 37.9#

58 71.4 56.4 84.6



#50

Toluene-d8

Concen: 48.465 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084769.D




Acq: 11 Nov 2024 14:06


Tgt Ion: 98 Resp: 333596

Ion Ratio Lower Upper

98 100

100 64.7 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 93.887 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 210440
Ion Ratio Lower Upper
43 100
58 41.0 31.6 47.4



#52
Toluene
Concen: 18.675 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 92 Resp: 90899
Ion Ratio Lower Upper
92 100
91 171.5 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 17.373 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 52220
Ion Ratio Lower Upper
75 100
77 31.9 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 17.631 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 75 Resp: 56090
Ion Ratio Lower Upper
75 100
77 31.0 24.2 36.2
39 44.3 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 18.690 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.006 min


Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion: 97 Resp: 34267




Ion Ratio Lower Upper

97 100

83 81.1 71.0 106.4

85 54.2 41.7 62.5

99 63.8 50.1 75.1



#56

Ethyl methacrylate

Concen: 19.635 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06




Tgt Ion: 69 Resp: 52072


Ion Ratio Lower Upper

69 100

41 58.4 53.3 79.9

39 28.6 28.8 43.2#









#57
1,3-Dichloropropane
Concen: 18.480 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 58151
Ion Ratio Lower Upper
76 100
78 32.8 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 86.954 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 63 Resp: 107908
Ion Ratio Lower Upper
63 100
106 29.6 21.8 32.6









#59
2-Hexanone
Concen: 95.560 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 155905
Ion Ratio Lower Upper
43 100
58 56.3 27.3 81.8



#60
Dibromochloromethane
Concen: 19.570 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 129 Resp: 40894
Ion Ratio Lower Upper
129 100
127 79.1 38.3 114.9









#61
1,2-Dibromoethane
Concen: 18.170 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 34114
Ion Ratio Lower Upper
107 100
109 94.1 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 50.214 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 95 Resp: 129173
Ion Ratio Lower Upper
95 100
174 76.4 0.0 145.2
176 71.3 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 247853
Ion Ratio Lower Upper
117 100
82 54.4 47.2 70.8
119 30.7 25.4 38.0



#64
Tetrachloroethene
Concen: 17.811 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion:164 Resp: 29364
Ion Ratio Lower Upper
164 100
166 121.7 101.0 151.4
129 91.1 80.8 121.2
131 94.1 78.1 117.1









#65
Chlorobenzene
Concen: 17.305 ug/l
RT: 11.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 95953
Ion Ratio Lower Upper
112 100
114 32.8 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 19.179 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion:131 Resp: 36990
Ion Ratio Lower Upper
131 100
133 90.9 47.9 143.7
119 61.5 32.8 98.3









#67
Ethyl Benzene
Concen: 17.483 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.006 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 161814
Ion Ratio Lower Upper
91 100
106 29.9 24.5 36.7



#68
m/p-Xylenes
Concen: 35.919 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion:106 Resp: 125863
Ion Ratio Lower Upper
106 100
91 207.6 167.1 250.7





#69

o-Xylene

Concen: 19.057 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.006 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:106 Resp: 62468

Ion Ratio Lower Upper

106 100

91 218.0 110.2 330.6



#70

Styrene

Concen: 18.264 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06




Tgt Ion:104 Resp: 104440


Ion Ratio Lower Upper

104 100

78 51.7 42.7 64.1

103 54.1 43.6 65.4





#71

Bromoform

Concen: 18.825 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:173 Resp: 27320

Ion Ratio Lower Upper

173 100

175 50.6 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06




Tgt Ion:152 Resp: 129321


Ion Ratio Lower Upper

152 100

115 63.2 31.3 93.9


150 163.0 0.0 349.8








#73
Isopropylbenzene
Concen: 16.874 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:105 Resp: 152918
Ion Ratio Lower Upper
105 100
120 26.1 12.9 38.7



#74
N-ethyl acetate
Concen: 15.919 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 43 Resp: 61403
Ion Ratio Lower Upper
43 100
70 46.9 34.4 51.6
55 27.2 19.6 29.4
61 26.0 19.6 29.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 16.680 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 83 Resp: 50456
 Ion Ratio Lower Upper
 83 100
 131 9.8 4.9 14.7
 85 64.4 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 25.245 ug/l
 RT: 12.988 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06

Tgt Ion: 75 Resp: 65259
 Ion Ratio Lower Upper
 75 100
 77 129.5 96.9 290.7









#77
Bromobenzene
Concen: 16.378 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 39902
Ion Ratio Lower Upper
156 100
77 211.9 115.1 345.3
158 98.5 48.9 146.8



#78
n-propylbenzene
Concen: 16.633 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 91 Resp: 176633
Ion Ratio Lower Upper
91 100
120 22.5 10.8 32.4









#79
2-Chlorotoluene
Concen: 16.982 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 117842
Ion Ratio Lower Upper
91 100
126 31.8 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 17.887 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion:105 Resp: 134354
Ion Ratio Lower Upper
105 100
120 48.1 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 17.943 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 19551
Ion Ratio Lower Upper
75 100
53 92.6 78.0 117.0
89 41.5 35.5 53.3



#82
4-Chlorotoluene
Concen: 16.040 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Tgt Ion: 91 Resp: 117134
Ion Ratio Lower Upper
91 100
126 32.3 15.7 47.1





#83

tert-Butylbenzene

Concen: 17.781 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:119 Resp: 110533

Ion Ratio Lower Upper

119 100

91 70.0 34.1 102.3

134 25.0 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 17.499 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN084769.D




Acq: 11 Nov 2024 14:06


Tgt Ion:105 Resp: 135654

Ion Ratio Lower Upper

105 100

120 45.1 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 16.909 ug/l
 RT: 13.617 min Scan# 1982
 Delta R.T. -0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:105 Resp: 148482
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 17.606 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06

Tgt Ion:119 Resp: 126790
 Ion Ratio Lower Upper
 119 100
 134 26.7 12.6 37.8
 91 25.8 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 15.646 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:146 Resp: 74392
 Ion Ratio Lower Upper
 146 100
 111 42.3 21.8 65.3
 148 64.1 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 16.436 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06


Tgt Ion:146 Resp: 74879
 Ion Ratio Lower Upper
 146 100
 111 42.6 21.2 63.6
 148 63.3 31.9 95.9








#89
n-Butylbenzene
Concen: 15.189 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 91 Resp: 102319
Ion Ratio Lower Upper
91 100
92 51.1 26.6 79.7
134 25.0 12.3 36.9



#90
Hexachloroethane
Concen: 16.171 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06


Tgt Ion: 117 Resp: 25970
Ion Ratio Lower Upper
117 100
201 75.2 34.4 103.2








#91
1,2-Dichlorobenzene
Concen: 16.038 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:146 Resp: 73274
Ion Ratio Lower Upper
146 100
111 43.9 22.0 66.0
148 64.3 32.1 96.3




#92
1,2-Dibromo-3-Chloropropane
Concen: 16.845 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN084769.D
Acq: 11 Nov 2024 14:06



Tgt Ion: 75 Resp: 10323
Ion Ratio Lower Upper
75 100
155 71.4 38.1 114.5
157 88.5 49.6 149.0









#93
 1,2,4-Trichlorobenzene
 Concen: 14.206 ug/l
 RT: 15.394 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:180 Resp: 35525
 Ion Ratio Lower Upper
 180 100
 182 88.4 47.0 141.0
 145 30.9 16.4 49.4



#94
 Hexachlorobutadiene
 Concen: 15.508 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. -0.000 min
 Lab File: VN084769.D
 Acq: 11 Nov 2024 14:06

Tgt Ion:225 Resp: 17031
 Ion Ratio Lower Upper
 225 100
 223 62.1 32.4 97.2
 227 62.5 32.4 97.0





#95

Naphthalene

Concen: 13.626 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:128 Resp: 101990

Ion Ratio Lower Upper

128 100

127 12.7 10.5 15.7

129 10.8 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 13.998 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN084769.D

Acq: 11 Nov 2024 14:06

Tgt Ion:180 Resp: 33980

Ion Ratio Lower Upper

180 100

182 94.4 48.2 144.6

145 34.4 17.3 51.9

