

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084767.D
 Acq On : 11 Nov 2024 13:18
 Operator : JC\MD
 Sample : P4774-06DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190-GW-638-640DL

Quant Time: Nov 11 13:36:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	263580	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	234257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	103754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107441	48.144	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.280%		
35) Dibromofluoromethane	8.165	113	88434	49.567	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.140%		
50) Toluene-d8	10.565	98	310695	47.275	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.540%		
62) 4-Bromofluorobenzene	12.847	95	115169	46.889	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	93.780%		
Target Compounds						
44) Trichloroethene	9.353	130	63076	34.419	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
Data File : VN084767.D
Acq On : 11 Nov 2024 13:18
Operator : JC\MD
Sample : P4774-06DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-GW-638-640DL

Quant Time: Nov 11 13:36:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

