

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084772.D
 Acq On : 11 Nov 2024 15:18
 Operator : JC\MD
 Sample : P4722-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1(5.5)

Quant Time: Nov 12 00:32:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	289915	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	269271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	115191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	119024	49.350	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.700%		
35) Dibromofluoromethane	8.159	113	95357	48.593	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.180%		
50) Toluene-d8	10.565	98	349810	48.391	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.780%		
62) 4-Bromofluorobenzene	12.847	95	131680	48.741	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.480%		
Target Compounds						
16) Acetone	4.436	43	17224	22.123	ug/l	94
20) Methylene Chloride	5.265	84	3558	1.677	ug/l #	73
25) 2-Butanone	7.482	43	10075	8.954	ug/l	90
43) Isopropyl Acetate	8.688	43	86312	18.309	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
Data File : VN084772.D
Acq On : 11 Nov 2024 15:18
Operator : JC\MD
Sample : P4722-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1(5.5)

Quant Time: Nov 12 00:32:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

