

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084774.D
 Acq On : 11 Nov 2024 16:06
 Operator : JC\MD
 Sample : P4722-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-3(5)

Quant Time: Nov 12 00:33:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	169890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	299832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126299	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	120968	49.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.600%
35) Dibromofluoromethane	8.165	113	97759	48.169	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.340%
50) Toluene-d8	10.565	98	358774	47.990	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.980%
62) 4-Bromofluorobenzene	12.847	95	134812	48.250	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.500%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	22989	29.922	ug/l	92
20) Methylene Chloride	5.259	84	3066	1.421	ug/l #	75
25) 2-Butanone	7.483	43	12665	11.063	ug/l	90
43) Isopropyl Acetate	8.688	43	94961	19.540	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
Data File : VN084774.D
Acq On : 11 Nov 2024 16:06
Operator : JC\MD
Sample : P4722-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-3(5)

Quant Time: Nov 12 00:33:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

