

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084781.D
 Acq On : 11 Nov 2024 18:54
 Operator : JC\MD
 Sample : P4792-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OILY-STONE-DRUM

Quant Time: Nov 12 00:37:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	278924	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	120720	49.263	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.520%
35) Dibromofluoromethane	8.159	113	98602	48.172	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.340%
50) Toluene-d8	10.565	98	361128	47.895	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.800%
62) 4-Bromofluorobenzene	12.847	95	137570	48.819	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.640%
Target Compounds						
16) Acetone	4.436	43	23989	31.395	ug/l	99
20) Methylene Chloride	5.259	84	4656	2.160	ug/l #	87
43) Isopropyl Acetate	8.688	43	207922	43.574	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
Data File : VN084781.D
Acq On : 11 Nov 2024 18:54
Operator : JC\MD
Sample : P4792-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OILY-STONE-DRUM

Quant Time: Nov 12 00:37:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

