

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084782.D
 Acq On : 11 Nov 2024 19:18
 Operator : JC\MD
 Sample : P4793-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M00-24-00345

Quant Time: Nov 12 00:37:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	162487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288820	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	132322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	115688	49.295	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	98.580%	
35) Dibromofluoromethane	8.159	113	93062	47.603	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.200%	
50) Toluene-d8	10.565	98	344458	47.832	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	95.660%	
62) 4-Bromofluorobenzene	12.847	95	142667	53.008	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	106.020%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	890660	1326.092	ug/l	96
17) Carbon Disulfide	4.694	76	8646	1.517	ug/l #	84
20) Methylene Chloride	5.271	84	12292	5.955	ug/l #	85
43) Isopropyl Acetate	6.688	43	156036	34.026	ug/l #	72
48) Methyl methacrylate	9.682	41	48880	25.531	ug/l	95
51) 4-Methyl-2-Pentanone	10.447	43	21200	9.040	ug/l	98
52) Toluene	10.629	92	2057682	404.069	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
Data File : VN084782.D
Acq On : 11 Nov 2024 19:18
Operator : JC\MD
Sample : P4793-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M00-24-00345

Quant Time: Nov 12 00:37:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

