

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111125\
 Data File : VN088254.D
 Acq On : 11 Nov 2025 14:04
 Operator : JC\MD
 Sample : Q3598-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-CONCRETE

Quant Time: Nov 12 01:15:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	203293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	455835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	446615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	210269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	200071	56.913	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.820%
35) Dibromofluoromethane	8.147	113	127192	41.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.060%
50) Toluene-d8	10.541	98	547363	46.388	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.780%
62) 4-Bromofluorobenzene	12.823	95	211483	50.749	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.500%
Target Compounds						
						Qvalue
16) Acetone	4.418	43	796435	450.965	ug/l	99
20) Methylene Chloride	5.265	84	13563	4.575	ug/l	90
25) 2-Butanone	7.477	43	23452	10.076	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111125\
Data File : VN088254.D
Acq On : 11 Nov 2025 14:04
Operator : JC\MD
Sample : Q3598-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-CONCRETE

Quant Time: Nov 12 01:15:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

