

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111125\
 Data File : VN088261.D
 Acq On : 11 Nov 2025 16:29
 Operator : JC\MD
 Sample : Q3581-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-240

Quant Time: Nov 12 01:18:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/12/2025
 Supervised By : Mahesh Dadoda 11/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	203051	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	456642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	464568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	220849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	204125	58.136	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	116.280%	
35) Dibromofluoromethane	8.147	113	151580	49.411	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	98.820%	
50) Toluene-d8	10.547	98	549651	46.500	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	93.000%	
62) 4-Bromofluorobenzene	12.829	95	212953	51.012	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	102.020%	
Target Compounds						Qvalue
16) Acetone	4.424	43	161103	91.330	ug/l	93
20) Methylene Chloride	5.253	84	13313m	4.496	ug/l	
25) 2-Butanone	7.477	43	15506	6.670	ug/l	91
43) Isopropyl Acetate	8.671	43	26742	3.011	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111125\
Data File : VN088261.D
Acq On : 11 Nov 2025 16:29
Operator : JC\MD
Sample : Q3581-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-240

Quant Time: Nov 12 01:18:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/12/2025
Supervised By :Mahesh Dadoda 11/12/2025

