

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111218\
 Data File : VN052257.D
 Acq On : 12 Nov 2018 12:58
 Operator : MD\SY
 Sample : PB114678ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114678ZHE#01

Quant Time: Nov 13 00:57:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	358933	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	590979	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	509791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	176337	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	229958	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
35) Dibromofluoromethane	7.59	113	186029	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
50) Toluene-d8	10.09	98	743758	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
62) 4-Bromofluorobenzene	12.40	95	214981	44.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.74%	
Target Compounds						
20) Methylene Chloride	4.55	84	8784	1.963	ug/l	89
43) Isopropyl Acetate	8.17	43	23463	3.000	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111218\
Data File : VN052257.D
Acq On : 12 Nov 2018 12:58
Operator : MD\SY
Sample : PB114678ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114678ZHE#01

Quant Time: Nov 13 00:57:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

