

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111218\
 Data File : VN052262.D
 Acq On : 12 Nov 2018 15:11
 Operator : MD\SY
 Sample : PB114678ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114678ZHE#06

Manual Integrations
 APPROVED

MMDadoda
 11/13/2018 2:54:35 PM

Quant Time: Nov 13 01:09:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	321666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	538402	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	472625	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	174433	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	213397	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
35) Dibromofluoromethane	7.59	113	169179	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
50) Toluene-d8	10.09	98	675068	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	12.40	95	205431	46.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.08%	
Target Compounds						
20) Methylene Chloride	4.56	84	9056m	2.258	ug/l	
43) Isopropyl Acetate	8.17	43	25359	3.559	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111218\
Data File : VN052262.D
Acq On : 12 Nov 2018 15:11
Operator : MD\SY
Sample : PB114678ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114678ZHE#06

Manual Integrations
APPROVED

MMDadoda
11/13/2018 2:54:35 PM

Quant Time: Nov 13 01:09:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

