

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111218\
 Data File : VN052267.D
 Acq On : 12 Nov 2018 17:25
 Operator : MD\SY
 Sample : PB114678ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114678ZHE#11

Quant Time: Nov 13 01:23:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	327450	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	549946	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	474555	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	167290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	218813	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	7.59	113	170729	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	10.09	98	691388	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	12.40	95	200765	44.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.06%	
Target Compounds						
20) Methylene Chloride	4.56	84	5232	1.282	ug/l	87
43) Isopropyl Acetate	8.17	43	25152	3.456	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111218\
Data File : VN052267.D
Acq On : 12 Nov 2018 17:25
Operator : MD\SY
Sample : PB114678ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114678ZHE#11

Quant Time: Nov 13 01:23:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

