

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111218\
 Data File : VN052269.D
 Acq On : 12 Nov 2018 18:18
 Operator : MD\SY
 Sample : PB114678ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114678ZHE#13

Quant Time: Nov 13 01:26:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	327533	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	554487	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	467211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	161259	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	221097	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
35) Dibromofluoromethane	7.59	113	173772	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	10.09	98	691953	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
62) 4-Bromofluorobenzene	12.40	95	196527	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
Target Compounds						
20) Methylene Chloride	4.56	84	6500	1.592	ug/l	86
43) Isopropyl Acetate	8.17	43	24847	3.386	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111218\
Data File : VN052269.D
Acq On : 12 Nov 2018 18:18
Operator : MD\SY
Sample : PB114678ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114678ZHE#13

Quant Time: Nov 13 01:26:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

