

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111218\
 Data File : VN052271.D
 Acq On : 12 Nov 2018 19:11
 Operator : MD\SY
 Sample : PB114678ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114678ZHE#15

Quant Time: Nov 13 01:30:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	325507	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	555643	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	473936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	160710	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	219984	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
35) Dibromofluoromethane	7.59	113	174418	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
50) Toluene-d8	10.09	98	689422	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
62) 4-Bromofluorobenzene	12.40	95	196773	43.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.40%	
Target Compounds						
20) Methylene Chloride	4.55	84	5721	1.410	ug/l #	86
43) Isopropyl Acetate	8.17	43	25164	3.422	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111218\
Data File : VN052271.D
Acq On : 12 Nov 2018 19:11
Operator : MD\SY
Sample : PB114678ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114678ZHE#15

Quant Time: Nov 13 01:30:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

