

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111218\  
 Data File : VN052272.D  
 Acq On : 12 Nov 2018 19:38  
 Operator : MD\SY  
 Sample : PB114678ZHE#16  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 23 Sample Multiplier: 28

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB114678ZHE#16

Quant Time: Nov 13 01:32:00 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N110918W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Nov 12 00:34:27 2018  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	347652	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	582132	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	494558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	167758	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	227575	47.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.72%	
35) Dibromofluoromethane	7.59	113	183378	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
50) Toluene-d8	10.09	98	728273	49.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.70%	
62) 4-Bromofluorobenzene	12.40	95	206219	43.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.42%	
Target Compounds						
20) Methylene Chloride	4.55	84	5638	1.301	ug/l	86
43) Isopropyl Acetate	8.17	43	23648	3.070	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111218\  
 Data File : VN052272.D  
 Acq On : 12 Nov 2018 19:38  
 Operator : MD\SY  
 Sample : PB114678ZHE#16  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 23 Sample Multiplier: 28

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB114678ZHE#16

Quant Time: Nov 13 01:32:00 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N110918W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Nov 12 00:34:27 2018  
 Response via : Initial Calibration

