


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111218\
Data File : VN052250.D
Acq On : 12 Nov 2018 9:03
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Title : SW846 8260
Last Update : Mon Nov 12 00:34:27 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.7	14204	PASS
75	95	30	60	51.4	32189	PASS
95	95	100	100	100.0	62608	PASS
96	95	5	9	6.8	4255	PASS
173	174	0.00	2	1.0	485	PASS
174	95	50	100	76.8	48077	PASS
175	174	5	9	8.0	3859	PASS
176	174	95	101	97.3	46762	PASS
177	176	5	9	5.9	2743	PASS