

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111219\
 Data File : VN059183.D
 Acq On : 12 Nov 2019 12:32
 Operator : JC/SP
 Sample : PB124708TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124708TB

Quant Time: Nov 13 06:11:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	406443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	641865	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	585057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	202136	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	273125	54.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.34%	
35) Dibromofluoromethane	7.57	113	205353	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
50) Toluene-d8	10.08	98	815305	54.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.02%	
62) 4-Bromofluorobenzene	12.40	95	251067	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
Target Compounds						
16) Acetone	3.80	43	14614	5.625	ug/l	93
18) Methyl Acetate	4.31	43	6283	1.019	ug/l #	82
43) Isopropyl Acetate	8.15	43	132297	12.433	ug/l #	91
80) 1,3,5-Trimethylbenzene	12.73	105	13457	1.080	ug/l	97
95) Naphthalene	15.13	128	10319	3.970	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111219\
Data File : VN059183.D
Acq On : 12 Nov 2019 12:32
Operator : JC/SP
Sample : PB124708TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124708TB

Quant Time: Nov 13 06:11:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

