

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111219\
 Data File : VN059190.D
 Acq On : 12 Nov 2019 15:07
 Operator : JC/SP
 Sample : K5799-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-303-SRI4

Quant Time: Nov 13 06:47:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	414896	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	640590	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	554422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	196274	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	276603	54.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.46%	
35) Dibromofluoromethane	7.57	113	210291	54.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.96%	
50) Toluene-d8	10.08	98	809033	53.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.40%	
62) 4-Bromofluorobenzene	12.40	95	254080	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
Target Compounds						
12) 1,1-Dichloroethene	3.72	96	4150	1.072	ug/l	89
16) Acetone	3.79	43	4479	1.689	ug/l	89
44) Trichloroethene	8.81	130	7061	1.593	ug/l	95
64) Tetrachloroethene	10.63	164	33868	8.984	ug/l	98
95) Naphthalene	15.13	128	48632	7.535	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

