

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064578.D
 Acq On : 12 Nov 2020 13:59
 Operator : JC/MD
 Sample : L4701-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRANSFORMER-1-1-TOP

Manual Integrations
 APPROVED

MMDadoda
 11/13/2020 11:42:28 AM

Quant Time: Nov 13 03:32:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	242875	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	527018	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	565984	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	201865	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	246577	58.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.58%	
35) Dibromofluoromethane	7.55	113	179833	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
50) Toluene-d8	10.06	98	688918	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	12.38	95	261149	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
Target Compounds						
16) Acetone	3.78	43	15897	9.865	ug/l #	88
18) Methyl Acetate	4.29	43	5960m	1.881	ug/l	
20) Methylene Chloride	4.51	84	9743m	3.047	ug/l	
43) Isopropyl Acetate	8.13	43	101752	11.277	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064578.D
Acq On : 12 Nov 2020 13:59
Operator : JC/MD
Sample : L4701-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRANSFORMER-1-1-TOP

Manual Integrations
APPROVED

MMDadoda
11/13/2020 11:42:28 AM

Quant Time: Nov 13 03:32:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

