

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064579.D
 Acq On : 12 Nov 2020 14:23
 Operator : JC/MD
 Sample : L4701-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRANSFORMER-1-1-BOTTOM

Quant Time: Nov 13 03:33:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	244799	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	528287	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	564353	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	204697	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	241530	56.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.28%	
35) Dibromofluoromethane	7.55	113	177631	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
50) Toluene-d8	10.06	98	694701	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	12.38	95	256748	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
Target Compounds						
16) Acetone	3.78	43	23980	14.764	ug/l	97
18) Methyl Acetate	4.29	43	6458	2.022	ug/l #	83
20) Methylene Chloride	4.50	84	10006	3.105	ug/l	99
43) Isopropyl Acetate	8.13	43	98254	10.863	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064579.D
Acq On : 12 Nov 2020 14:23
Operator : JC/MD
Sample : L4701-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRANSFORMER-1-1-BOTTOM

Quant Time: Nov 13 03:33:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

