

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064580.D
 Acq On : 12 Nov 2020 14:47
 Operator : JC/MD
 Sample : L4701-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRANSFORMER-1-2-TOP

Manual Integrations
 APPROVED

MMDadoda
 11/13/2020 11:42:31 AM

Quant Time: Nov 13 03:35:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	229016	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	510189	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	545590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	193069	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	236639	59.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.64%	
35) Dibromofluoromethane	7.55	113	171754	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	10.06	98	657888	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	12.38	95	246848	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
Target Compounds						
16) Acetone	3.79	43	15288	10.061	ug/l	92
18) Methyl Acetate	4.29	43	5986m	2.004	ug/l	
20) Methylene Chloride	4.51	84	10342	3.430	ug/l	87
43) Isopropyl Acetate	8.13	43	98098	11.231	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064580.D
Acq On : 12 Nov 2020 14:47
Operator : JC/MD
Sample : L4701-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRANSFORMER-1-2-TOP

Manual Integrations
APPROVED

MMDadoda
11/13/2020 11:42:31 AM

Quant Time: Nov 13 03:35:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

