

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064584.D
 Acq On : 12 Nov 2020 16:25
 Operator : JC/MD
 Sample : L4701-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRANSFORMER-1-4-TOP

Quant Time: Nov 13 03:41:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	225607	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	499499	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	540205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	196675	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	238258	60.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.26%	
35) Dibromofluoromethane	7.55	113	171450	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
50) Toluene-d8	10.06	98	651122	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	12.38	95	244116	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
Target Compounds						
16) Acetone	3.78	43	87982	58.776	ug/l	98
18) Methyl Acetate	4.29	43	6382	2.169	ug/l #	61
20) Methylene Chloride	4.51	84	8446	2.844	ug/l	95
43) Isopropyl Acetate	8.13	43	98664	11.537	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064584.D
Acq On : 12 Nov 2020 16:25
Operator : JC/MD
Sample : L4701-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRANSFORMER-1-4-TOP

Quant Time: Nov 13 03:41:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

