

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064589.D
 Acq On : 12 Nov 2020 18:27
 Operator : JC/MD
 Sample : L4701-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRANSFORMER-5-2-BOTTOM

Quant Time: Nov 13 03:51:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	228468	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	493901	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	525160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	179697	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235865	59.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.54%	
35) Dibromofluoromethane	7.55	113	173975	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
50) Toluene-d8	10.06	98	648851	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
62) 4-Bromofluorobenzene	12.38	95	234237	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
Target Compounds						
16) Acetone	3.79	43	14550	9.598	ug/l #	81
18) Methyl Acetate	4.29	43	6386	2.143	ug/l	98
20) Methylene Chloride	4.50	84	9463	3.146	ug/l #	88
43) Isopropyl Acetate	8.13	43	96891	11.459	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064589.D
Acq On : 12 Nov 2020 18:27
Operator : JC/MD
Sample : L4701-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRANSFORMER-5-2-BOTTOM

Quant Time: Nov 13 03:51:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

