

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064591.D
 Acq On : 12 Nov 2020 19:15
 Operator : JC/MD
 Sample : L4701-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRANSFORMER-5-3-BOTTOM

Quant Time: Nov 13 03:54:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226548	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	496672	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	538076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	190737	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235540	59.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.38%	
35) Dibromofluoromethane	7.55	113	174060	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
50) Toluene-d8	10.06	98	657531	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
62) 4-Bromofluorobenzene	12.38	95	245875	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
Target Compounds						
16) Acetone	3.79	43	14948	9.944	ug/l	92
18) Methyl Acetate	4.29	43	7274	2.461	ug/l #	88
20) Methylene Chloride	4.50	84	9484	3.180	ug/l #	69
43) Isopropyl Acetate	8.13	43	100077	11.769	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064591.D
Acq On : 12 Nov 2020 19:15
Operator : JC/MD
Sample : L4701-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRANSFORMER-5-3-BOTTOM

Quant Time: Nov 13 03:54:33 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

