

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064606.D
 Acq On : 13 Nov 2020 2:32
 Operator : JC/MD
 Sample : PB132945ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132945ZHE#04

Quant Time: Nov 13 04:37:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	234446	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	517173	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	549195	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	198517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	233996	57.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.60%	
35) Dibromofluoromethane	7.55	113	171593	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
50) Toluene-d8	10.06	98	664994	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	12.38	95	250924	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
Target Compounds						
16) Acetone	3.79	43	13580	8.730	ug/l	91
18) Methyl Acetate	4.29	43	6018	1.968	ug/l #	86
20) Methylene Chloride	4.50	84	7160	2.320	ug/l	87
43) Isopropyl Acetate	8.13	43	98421	11.116	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064606.D
Acq On : 13 Nov 2020 2:32
Operator : JC/MD
Sample : PB132945ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132945ZHE#04

Quant Time: Nov 13 04:37:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

