

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064607.D
 Acq On : 13 Nov 2020 2:56
 Operator : JC/MD
 Sample : PB132945ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132945ZHE#05

Quant Time: Nov 13 04:38:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	220155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	491790	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	526563	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	184360	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	227919	59.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.88%	
35) Dibromofluoromethane	7.55	113	166492	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
50) Toluene-d8	10.06	98	640328	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
62) 4-Bromofluorobenzene	12.38	95	240276	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
Target Compounds						
16) Acetone	3.79	43	13797	9.445	ug/l	95
18) Methyl Acetate	4.29	43	5729	1.995	ug/l #	89
20) Methylene Chloride	4.50	84	7946	2.742	ug/l	96
43) Isopropyl Acetate	8.13	43	100213	11.902	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064607.D
Acq On : 13 Nov 2020 2:56
Operator : JC/MD
Sample : PB132945ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132945ZHE#05

Quant Time: Nov 13 04:38:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

