

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064615.D
 Acq On : 13 Nov 2020 6:10
 Operator : JC/MD
 Sample : L4622-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B12-110320-16-17

Manual Integrations
 APPROVED

MMDadoda
 11/13/2020 11:43:54 AM

Quant Time: Nov 13 07:43:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	220056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	499311	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	522024	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	184632	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	230707	60.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.38%	
35) Dibromofluoromethane	7.55	113	166376	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
50) Toluene-d8	10.06	98	637291	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
62) 4-Bromofluorobenzene	12.38	95	232958	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
Target Compounds						
16) Acetone	3.78	43	19989m	13.690	ug/l	Qvalue
18) Methyl Acetate	4.30	43	5492	1.913	ug/l #	86
20) Methylene Chloride	4.50	84	7233	2.497	ug/l	94
43) Isopropyl Acetate	8.13	43	94535	11.059	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064615.D
Acq On : 13 Nov 2020 6:10
Operator : JC/MD
Sample : L4622-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B12-110320-16-17

Manual Integrations
APPROVED

MMDadoda
11/13/2020 11:43:54 AM

Quant Time: Nov 13 07:43:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

