

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064616.D
 Acq On : 13 Nov 2020 6:34
 Operator : JC/MD
 Sample : L4631-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B11-110420-0-1

Quant Time: Nov 13 07:45:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	221038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	505511	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	529252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191443	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	229247	59.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.08%	
35) Dibromofluoromethane	7.55	113	170089	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
50) Toluene-d8	10.06	98	643709	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
62) 4-Bromofluorobenzene	12.38	95	239720	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
Target Compounds						
16) Acetone	3.78	43	13604	9.276	ug/l	95
18) Methyl Acetate	4.29	43	4827	1.674	ug/l	96
20) Methylene Chloride	4.51	84	10710	3.681	ug/l	97
43) Isopropyl Acetate	8.13	43	94731	10.946	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064616.D
Acq On : 13 Nov 2020 6:34
Operator : JC/MD
Sample : L4631-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B11-110420-0-1

Quant Time: Nov 13 07:45:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

