

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
 Data File : VN064619.D
 Acq On : 13 Nov 2020 7:47
 Operator : JC/MD
 Sample : L4705-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1

Manual Integrations
 APPROVED

MMDadoda
 11/13/2020 11:44:16 AM

Quant Time: Nov 13 08:22:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	231630	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	517047	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	551410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	210300	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235382	58.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.68%	
35) Dibromofluoromethane	7.55	113	172188	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	10.06	98	661141	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	12.38	95	254577	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
Target Compounds						
16) Acetone	3.78	43	26053	16.952	ug/l	98
18) Methyl Acetate	4.29	43	6105	2.021	ug/l #	82
20) Methylene Chloride	4.51	84	10092m	3.310	ug/l	
43) Isopropyl Acetate	8.13	43	95925	10.836	ug/l	98
52) Toluene	10.12	92	16923	1.729	ug/l	93
95) Naphthalene	15.11	128	18525	5.113	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111220\
Data File : VN064619.D
Acq On : 13 Nov 2020 7:47
Operator : JC/MD
Sample : L4705-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1

Manual Integrations
APPROVED

MMDadoda
11/13/2020 11:44:16 AM

Quant Time: Nov 13 08:22:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

