

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111221\
 Data File : VN069467.D
 Acq On : 12 Nov 2021 14:48
 Operator : JC/MD
 Sample : M4563-01
 Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-1

Quant Time: Nov 15 03:26:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1198871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2116383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2038226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	974579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	901670	51.871	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.740%	
35) Dibromofluoromethane	8.019	113	596970	47.031	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 94.060%	
50) Toluene-d8	10.440	98	2694492	53.974	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.940%	
62) 4-Bromofluorobenzene	12.734	95	976650	53.254	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.500%	
Target Compounds						
					Qvalue	
59) 2-Hexanone	11.092	43	24305	9.058	ug/l	99
70) Styrene	12.299	104	4935	1.624	ug/l	94
85) sec-Butylbenzene	13.503	105	32516	0.423	ug/l	92
86) p-Isopropyltoluene	13.616	119	19064	0.313	ug/l	93
95) Naphthalene	15.507	128	126287	7.110	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111221\
Data File : VN069467.D
Acq On : 12 Nov 2021 14:48
Operator : JC/MD
Sample : M4563-01
Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-1

Quant Time: Nov 15 03:26:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

