

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111221\
 Data File : VN069471.D
 Acq On : 12 Nov 2021 16:41
 Operator : JC/MD
 Sample : M4675-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 15 03:28:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1240847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2188719	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2189500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	817494	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	926050	51.471	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.940%
35) Dibromofluoromethane	8.024	113	635701	48.427	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.860%
50) Toluene-d8	10.443	98	2834179	54.896	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.800%
62) 4-Bromofluorobenzene	12.733	95	995118	52.468	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111221\
Data File : VN069471.D
Acq On : 12 Nov 2021 16:41
Operator : JC/MD
Sample : M4675-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 15 03:28:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

