

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111221\
 Data File : VN069475.D
 Acq On : 12 Nov 2021 18:23
 Operator : JC/MD
 Sample : M4676-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SY-2R-20211111

Quant Time: Nov 15 03:30:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1192183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2100185	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2077724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	753637	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	900880	52.116	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.240%
35) Dibromofluoromethane	8.024	113	615368	48.854	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.700%
50) Toluene-d8	10.443	98	2685051	54.200	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.400%
62) 4-Bromofluorobenzene	12.733	95	928293	51.008	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.020%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	16028	1.719	ug/l	# 85
52) Toluene	10.505	92	22750	0.669	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111221\
Data File : VN069475.D
Acq On : 12 Nov 2021 18:23
Operator : JC/MD
Sample : M4676-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SY-2R-20211111

Quant Time: Nov 15 03:30:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

