

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
 Data File : VN084793.D
 Acq On : 12 Nov 2024 14:01
 Operator : JC\MD
 Sample : PB164761TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164761TB

Quant Time: Nov 12 23:35:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	162036	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	277946	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237845	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	103613	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112433	48.041	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	96.080%	
35) Dibromofluoromethane	8.159	113	90740	48.231	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.460%	
50) Toluene-d8	10.565	98	316292	45.639	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	91.280%	
62) 4-Bromofluorobenzene	12.847	95	114641	44.261	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.520%	
Target Compounds						
20) Methylene Chloride	5.271	84	12552	6.098	ug/l	85
93) 1,2,4-Trichlorobenzene	15.388	180	2591	1.293	ug/l	98
95) Naphthalene	15.641	128	6153	1.026	ug/l #	94
96) 1,2,3-Trichlorobenzene	15.835	180	2090	1.075	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
Data File : VN084793.D
Acq On : 12 Nov 2024 14:01
Operator : JC\MD
Sample : PB164761TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164761TB

Quant Time: Nov 12 23:35:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

