

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
 Data File : VN084794.D
 Acq On : 12 Nov 2024 14:25
 Operator : JC\MD
 Sample : P4722-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1(0-6)

Quant Time: Nov 12 23:36:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	177522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304861	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124231	48.452	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.900%		
35) Dibromofluoromethane	8.165	113	99722	48.326	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.660%		
50) Toluene-d8	10.565	98	343425	45.179	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.360%		
62) 4-Bromofluorobenzene	12.847	95	136989	48.220	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.440%		
Target Compounds						
16) Acetone	4.442	43	7647	7.558	ug/l #	75
20) Methylene Chloride	5.265	84	6630	2.940	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
Data File : VN084794.D
Acq On : 12 Nov 2024 14:25
Operator : JC\MD
Sample : P4722-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1(0-6)

Quant Time: Nov 12 23:36:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

