

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
 Data File : VN084800.D
 Acq On : 12 Nov 2024 16:49
 Operator : JC\MD
 Sample : P4799-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-TA2-03-G

Quant Time: Nov 12 23:39:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	170276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	297743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	121229	49.293	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.580%		
35) Dibromofluoromethane	8.165	113	83230	41.298	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	82.600%		
50) Toluene-d8	10.565	98	338922	45.653	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.300%		
62) 4-Bromofluorobenzene	12.847	95	137454	49.541	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.080%		
Target Compounds						
16) Acetone	4.430	43	88577	123.237	ug/l	97
17) Carbon Disulfide	4.700	76	6337	1.061	ug/l #	86
18) Methyl Acetate	5.024	43	13649	1.378	ug/l	99
20) Methylene Chloride	5.271	84	4416	2.042	ug/l #	71
95) Naphthalene	15.641	128	140999	19.586	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
Data File : VN084800.D
Acq On : 12 Nov 2024 16:49
Operator : JC\MD
Sample : P4799-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-TA2-03-G

Quant Time: Nov 12 23:39:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

