

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
 Data File : VN084804.D
 Acq On : 12 Nov 2024 18:26
 Operator : JC\MD
 Sample : P4807-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7

Quant Time: Nov 13 02:35:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	179344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	312429	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	136250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	120809	46.638	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	93.280%	
35) Dibromofluoromethane	8.165	113	100264	47.411	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.820%	
50) Toluene-d8	10.565	98	371188	47.649	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	95.300%	
62) 4-Bromofluorobenzene	12.847	95	149898	51.486	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	102.980%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	32980	41.700	ug/l #	85
18) Methyl Acetate	5.018	43	14751	1.527	ug/l	99
20) Methylene Chloride	5.259	84	5585	2.451	ug/l #	76
43) Isopropyl Acetate	8.688	43	104542	20.700	ug/l #	91
95) Naphthalene	15.635	128	231513	29.357	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
Data File : VN084804.D
Acq On : 12 Nov 2024 18:26
Operator : JC\MD
Sample : P4807-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7

Quant Time: Nov 13 02:35:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

